

# Inspection Certificate 3.1\*

Hans Schmidt & Co confirms that the measuring instrument, which is referred below, was manufactured according to our technical specifications. The corresponding calibration report is enclosed.

Object: Tension Meter

Model: DTX-1000  
Serial No. 921-00156

Customer: XXX

Reference (Order No.): 1234

Date of calibration: 05.05.2020

Hans Schmidt & Co GmbH

Person in charge *T. Müller*  
T. Müller

Quality Manager (QA)  
R. Meier

We suggest a recalibration period of 1 year, depending on the usage of the instrument. The most favorable period between calibrations has to be defined by quality-assuring personal, corresponding to the operating time of the instrument.

\* For tension meters are no international standards available. Therefor we use the definition of quality checks and verification certificates as described in DIN EN 10204 section 3 to 6.

## Calibration Report

**Calibration:** According to SCHMIDT-Factory procedure No. 02

Model: DTX-1000 .....

Temperature: 22 °C

☒ Standard unit ☐ Customized

Measuring Range 10 - 1000 .....

Serial Number: 921-00156 .....

Reading in ☒ cN ☐ daN ☐ .....

|    | Test Weight | Position 1 | Position 2 | Position | Position |
|----|-------------|------------|------------|----------|----------|
| 1. | 10          | 11         | 11         |          |          |
| 2. | 100         | 101        | 98         |          |          |
| 3. | 200         | 198        | 200        |          |          |
| 4. | 500         | 499        | 501        |          |          |
| 5. | 700         | 700        | 699        |          |          |
| 6. | 900         | 898        | 901        |          |          |
| 7. | 1000        | 1001       | 1000       |          |          |

**Accuracy:** For PA from 5 % to 100 % of range  $\pm 0.5$  % Full Scale and  $\pm 1$  Digit  
other tension ranges and calibration material  $\pm 3$  % Full Scale and  $\pm 1$  Digit

### Calibration material / Instrument settings

| Position | Calibration Material | Material Diameter | Material Thickness Compensator |
|----------|----------------------|-------------------|--------------------------------|
| 1        | Polyamide Monofil    | 0.3               | 0.3                            |
| 2        | Copper Wire          | 0.25              | 0.2 - 0.3                      |
|          |                      |                   |                                |
|          |                      |                   |                                |

**Verification:** ☒ Instrument working properly, calibration is within the tolerance limits.

**Traceability of test weights:** Hooked weights, accuracy class M3, calibration in Newton at falling speed of 9,80735 m/s<sup>2</sup> (location Waldkraiburg).

### Calibration test of weights:

Calibration of hooked weights:

Electronic balance, measuring tolerance to 4.8 kg  $\pm 0.9$  g, higher  $\pm 5$ g

Calibration of balance with analytical weights: 500-g weight, class F1, calibration test report no. G8-603, calibration mark: D-K-19408-01-00

2-kg weight, class F1, calibration test report no. G8-601, calibration mark: D-K-19408-01-00

10-kg weight, class F1, calibration test report no. G8-602, calibration mark: D-K-19408-01-00

Calibrated date: 05.05.2020

Calibration Engineer: John Smith